

DIACHECK IMMUNOASSAY CONTROL SET

LOT 24010101

 23.07.2026

Test Name	Level 1			Level 2		
	Result	SD	Unit	Result	SD	Unit
Estradiol	183.9	25	pg/mL	965	110	pg/mL
Ferritin	78.62	11	ng/mL	100	15	µg/L
Folik Asit	3.89	1.49	ng/mL	4.5	2	ng/mL
FSH	24.6	4.3	mIU/mL	3.44	1.4	mIU/mL
FT3	5.46	0.6	pmol/L	5.58	1.2	pmol/L
FT4	18.02	0.2	pmol/L	18.08	2.48	pmol/L
HCG	426	48	mIU/mL	1847	150	mIU/mL
İnsülin	10.05	1.96	µIU/mL	10.5	1.96	µIU/mL
LH	16.42	2.53	IU/L	8.76	0.5	IU/L
Progesteron	1.99	0.56	ng/mL	7.69	1.5	ng/mL
Prolaktin	325	32.9	mIU/L	650	42	mIU/L
PSA Total	0.03	0.02	ng/mL	0.1	0.04	ng/mL
PTH	3.09	2.1	pg/ml	19.5	3.99	pg/ml
Testosteron	1.65	0.4	ng/mL	1.65	0.4	ng/mL
TSH	2.15	0.32	µIU/mL	2.47	0.32	µIU/mL
Vitamin B12	300	48	pg/mL	270	35	pg/mL
Vitamin D	24	8.2	ng/mL	22	8.19	ng/mL

IVD 

Stability

-20 °C 18 Month

2-8 °C 14 Day, Open vial is 2-8 °C 10 Day



Medios Medikal Bilişim Danışmanlık Ticaret Limited Şirketi
Güzelevler Mahallesi Çınar Cad. No:5 Yüreğir/Adana